



Republic of the Philippines
GENERAL SANTOS CITY WATER DISTRICT
E. Fernandez Street, Barangay Lagao General Santos
Telephone No.: 552-3824; Telefax No.: 553-4960
Email Address: gscwdprocurement@yahoo.com

GPM-014-2

REQUEST FOR QUOTATION

REQ #: 3-0221-011 DATE: 02/03/2021

SUPPLIER: _____
ADDRESS: _____
TELEPHONE NO. : _____
TIN NO. : _____

Dear Sir/Madam,


Pls. quote your lowest price on the item/s listed below subject to general condition stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 02/09/2021, Pls. furnish this office the following articles subject to the terms and conditions contained therein.

ITEM NO.	DESCRIPTION	QTY.	UNIT	UNIT PRICE
1	IR CAMERA SCANNER : PLEASE SEE ATTACHED SPECIFICATIONS-ABC: PHP 130,000.00 ##### nothing follows #####	2	unit	P

Date of Delivery: _____
Price Validity : _____
Terms of Payment: _____

Place of Delivery: GSCWD Bodega
****No partial delivery allowed****
Note: 6% WITHHOLDING TAX DEDUCTION
PR NO.: 0121-085

Requested by:


GESEM R. CABAÑERO
MATERIAL PROCUREMENT
OFFICER A

Price quoted by:

Supplier

Date



TECHNICAL SPECIFICATION

TITLE: Purchase of IR Camera Scanner

1.0 SCOPE OF CONTRACT

The scope shall be supply and delivery of one (1) unit IR Camera Scanner to General Santos City Water District (GSCWD).

2.0 IR Camera Scanner Specification

Particular		Requirements
Infrared heat map overlay		five blending modes
Field of view		28° x 28°
High/low temperature alarms		With alarm
Battery type		Four (4) AA batteries
Ergonomics		Slim pocket-sized design
Temperature measurement range		-10 °C to +250 °C (14 °F to 482 °F)
Temperature measurement accuracy		± 2 °C or ± 2 %
Temperature measurement		center-point
Storage medium (micro SD card)		Stores up to 10,000 images per GB (4 GB card included)
Infrared spectral band		6.5 µm to 14 µm
Level and span		Auto
Focus mechanism		Focus free



Alignment of blended visual image and infrared heat map	NEAR: 23 cm (9 in) from target FAR: >23 cm (9 in) from target
Safety and compliance	CFR47: 2009 Class A. Part 15 subpart B; CE: EN 61326:2006; IEC/EN 61010-1:2010


3.0 TERMS AND CONDITIONS


1. The scope of work shall be supply and delivery of IR Camera Scanner to General Santos City Water District (GSCWD).
2. The winning bidder is required to submit actual documents (data sheets, manufacturer's test results, and etc.) of the units.
3. In case the unit/units failed to pass the minimum requirements during commissioning, the winning bidder shall replace the unit/units without cost to GSCWD.
4. Field testing shall be made once the units are ready for operation. GSCWD engineers shall conduct field testing with the presence of winning bidder's representative.
5. The winning bidder shall deliver the IR Camera Scanner within **Thirty (30)** calendar days upon receipt of Purchase Order (PO).
6. The winning bidder shall be subjected to Liquidated Damages (LD) for each day of delay as provided by the IRR of RA 9184.
7. That the winning bidder shall issue a warranty certificate of not less than one (1) year and the warranty shall commence from the date of commissioning of the pumping equipment.




8. That the bidder must have an available qualified and trained technicians 24/7 in case of unit/units failure.

Prepared by:



EDMUND L. BADAL, REE
TWG – Member


DLUANIE R. ENERIO, ME
TWG – Member


JUDY LIM PASKAN
TWG – Member


MICHAEL S. GABALES, REE
TWG – HEAD

Recommending Approval:


ROGELIO A. BESANA JR. C.E./RMP
Manager A – Engineering & Const'n. Dept.

Approval:

ARN B. GELLANGARIN, PME
General Manager A