



Republic of the Philippines
GENERAL SANTOS CITY WATER DISTRICT
E. Fernandez Street, Barangay Lagao General Santos
Telephone No.: 552-3824; Telefax No.: 553-4960
Email Address: gscwdprocurement@yahoo.com

GPM-014-2

REQUEST FOR QUOTATION

RFQ #: 3-0322-071 DATE: 03/15/2022

SUPPLIER: _____
ADDRESS: _____
TELEPHONE NO. : _____
TIN NO. : _____

Dear Sir/Madam,

Pls. quote your lowest price on the item/s listed below subject to general condition stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 03/21/2022, Pls. furnish this office the following articles subject to the terms and conditions contained therein.

ITEM NO.	DESCRIPTION	QTY.	UNIT	UNIT PRICE
1	CHLORINE TEST KIT : PLEASE SEE ATTACHED SPECIFICATIONS-ABC: PHP 200,000.00 ##### nothing follows #####	20	kit	P

Date of Delivery: _____
Price Validity : _____
Terms of Payment: _____


Place of Delivery: GSCWD Bodega

*****No partial delivery allowed****

Note: 6% WITHHOLDING TAX DEDUCTION

PR NO.: 0322-033

Requested by:


GESEM R. CABANERO
MATERIAL PROCUREMENT
OFFICER A

Price quoted by:

Supplier

Date



TECHNICAL SPECIFICATION

TITLE: PURCHASE OF CHLORINE TEST KITS

1.0 General Scope

The scope of work shall be supply and delivery of Chlorine Test Kits to General Santos City Water District.

2.0 General Specification

Item	Technical Specification	Statement of Compliance
1	Free Chlorine Medium Range Test Kit	
2	Free Chlorine measurement type: Checker Disk	
3	Free Chlorine Range: 0 to 3.5 mg/L	
4	Free Chlorine Resolution: 0.1mg/L	
5	Chlorine Method: DPD (Colorimetric Method)	
6	Required Materials: Glass Vials with caps, Plastic Pipette, Checker Disc.	




Republic of the Philippines
GENERAL SANTOS CITY WATER DISTRICT (GSCWD)
GSCWD Bldg., E. Fernandez Street (Emergency Road), Lagao, General Santos City
Telephone No.: 552-3824/554-7231; Telefax No.: 553-4960
E-mail Address: gscwaterdistrict@yahoo.com

3.0 TERMS AND CONDITIONS

1. The scope of work shall be supply and delivery of **Chlorine Test Kits** to General Santos City Water District (GSCWD).
2. The winning bidder is required to submit the following but not limited to data sheets and operating manual.
3. The winning bidder shall deliver the **Chlorine Test Kits** within **Thirty (30)** calendar days upon receipt of Purchase Order (PO).
4. The winning bidder shall be subjected to Liquidated Damages (LD) for each day of delay as provided by the IRR of RA 9184.
5. That the winning bidder shall issue a warranty certificate of not less than one (1) year.

Prepared by:


EDMUND L. BADAL, REE
OIC-Principal Engineer C