



Republic of the Philippines
GENERAL SANTOS CITY WATER DISTRICT
E. Fernandez Street, Barangay Lagao General Santos
Telephone No.: 552-3824; Telefax No.: 553-4960
Email Address: gscwdprocurement@yahoo.com

GPM-014-3

REQUEST FOR QUOTATION

RFQ #: 3-0323-099 DATE: 03/30/2023

SUPPLIER: _____
ADDRESS: _____
TELEPHONE NO. : _____
TIN NO. : _____

Dear Sir/Madam,

Pls. quote your lowest price on the item/s listed below subject to general condition stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 04/05/2023, Pls. furnish this office the following articles subject to the terms and conditions contained therein.

ITEM NO.	DESCRIPTION	QTY.	UNIT	UNIT PRICE
1	DISTILLING APPARATUS/WATER PURIFIER : SEE ATTACHED SPECIFICATIONS AND SAMPLE PHOTOS-ABC: PHP 300,000.00 ##### nothing follows #####	1	unit	P

Delivery term: _____
Price Validity : _____
Terms of Payment: _____

Place of Delivery: GSCWD Bodega
*****No partial delivery allowed****
Note: 6% WITHHOLDING TAX DEDUCTION
PR NO.: 0223-049


Requested by:

GESEM R. CADAÑERO
MATERIAL PROCUREMENT
OFFICE A

Price quoted by:

Supplier

Date



SPECIFICATIONS:

1. Inlet/Feed Water : Tap
2. Output : 10-30L/hour
3. Outlet/Produce : Distilled, Demineralized or Ultrapure water 2 (or combination)
4. Electrical Requirement : 220-240 V, 50/60 Hz
5. Flow rate : 2L per minute (with pressure tank)
6. Dimension and Weight: Approximately :
L x W x H
410 x 220 x 420mm
18-32 kg
7. Standard Configuration:
 - Main Body – includes 1 set of cartridge
 - 15-30L Pressure Tank
 - TDS/Conductivity Test Pen

Mari
Ma. Rita B. Barriases, RMT
PRC No. 0023789